



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE May 26, 2010 INITIALS: DEC OWNER: WASHINGTON STATE PATROL

DISTRICT: 3 ADDRESS: SR97 MP56

LOCATION Toppenish 42

COUNTY: Yakima

MAKE: Fairbanks

TYPE Trk Scale

ACCURACY CLASS: III L

MODEL: 42-9000-BW

ID. NO: G608891

CAPACITY: 62500 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 50 lbs.

REG. ELEMENT: Dial

PLATFORM 8x12

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	4000	4000	0	CENTER	20000	20000	0
CENTER	8000	8000	0	SEC 1	20000	20000	0
CENTER	12000	12000	0	SEC 2	20000	20000	0
CENTER	16000	16000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by:

David Cromer

Date:

June 28, 2010

Print name: David Cromer

State of Washington

County of Shurston

Signed and sworn to (or affirmed before
me on June 28, 2010 by

Brenda A Biscay

(Signature of notary public)

Brenda A Biscay

Printed Name of Notary

My appointment expires 3-12-12

